

Top five tips for Onyx battery care

Below are our top five tips that will ensure you get the most out of the innovative 24/7 Onyx battery and charger system.

TIP #1. Always use fully charged batteries

This will give you the maximum working time on the Venus before needing to change batteries.

TIP #2. Replace batteries one at a time

If both batteries are removed at the same time, the Venus will lose all power and turn off. This could harm your precious medical data.

TIP #3. Do not let the batteries be completely drained

This ensures the batteries stay in good health and extends overall battery life. When either light on the bottom of the Venus is purple, it's time to change the battery. If the purple light is flashing, this means the battery must be changed immediately, before continuing work.

TIP #4. Recharge batteries as soon as possible

Recharge your batteries as soon as possible after removing. This will extend the life span and performance of the battery.

TIP #5. Use 'hibernation' mode on your device

Talk to your IT support team to set up a hibernation mode for the Venus. This greatly extends time between battery changes.

For more information please contact us on 1300 723 282
www.peacockshealthcare.com

VIC | NSW | QLD | WA | SA | NZ